

Infection Control and Maintenance Guidelines



For Bedfont® GastroCH₄ECK® Gastrolyzer®

Before using any of the above equipment, the Operating Manual should be read and fully understood.

Infection Control

Washing hands before and after testing is highly recommended for both operator and user as part of a sensible infection control regime. The GastroCH₄ECK® Gastrolyzer® breath sample mouthpieces are specifically designed with the latest bacterial filtration to remove >99% of airborne bacteria and >98% of viruses, further removing any moisture from the patients breath.

GastroCH₄ECK® Gastrolyzer® breath bags are single-use only as re-use can increase the risk of cross-infection. One-way breath bag mouthpieces are supplied to ensure no loss of sample and increased infection control.

The GastroCH₄ECK® Gastrolyzer® device comes with a moisture removal filter. The filter will gradually change colour from orange (dry) to dark green (saturated), indicating that replacement is required.

Please note that these have a 3 month shelf life and will last for approximately 150 samples.

Routine maintenance

See the operating manual for full maintenance instructions.

The GastroCH₄ECK® Gastrolyzer® should be switched off once every 24 hours. The device must be calibrated once every 4 weeks. The device will provide daily reminders via the reminder screen after turning on the GastroCH₄ECK® Gastrolyzer® for 1 month leading up to the sensor change due date. Once this date has elapsed the date will turn red to indicate that sensor replacement is overdue:

- The H₂ sensor should be replaced every 2 years;
- The O₂ sensor should be replaced every 2 years;
- The CH₄ sensor should be replaced every 5 years.



Bedfont® Scientific Ltd.
Station Road, Harrietsham, Maidstone,
Kent, ME17 1JA England
Tel: +44 (0)1622 851122 Fax: +44 (0)1622 854860
Email: ask@bedfont.com Web: www.bedfont.com

© Bedfont® Scientific Limited 2024

Issue 5 - September 2024, Part No: LAB807_USA
Bedfont® Scientific Limited reserves the right to change or update this literature without prior notice.
Registered in: England and Wales. Registered No: 1289798



Infection Control and Maintenance Guidelines



Calibration

- The GastroCH₄ECK® Gastrolyzer® should be calibrated with 100 ppm H₂ & CH₄ once a month
- It is recommended that only Bedfont® calibration gas is used; the unit may give false readings if not calibrated.

Cleaning

Wipe the external surfaces of the instrument with a product specifically developed for this purpose. The device should be cleaned for initial use and after each patient use. Bedfont® recommends wiping the instrument external surfaces between each patient with an alcohol-free wipe specifically designed for this purpose. A list of suitable wipes can be found here: <https://www.bedfont.com/cleaning-bedfont-devices>.

NEVER use alcohol or cleaning products containing alcohol or other organic solvents as these vapours will damage the sensor within the instrument. Under no circumstances should the instrument be immersed in or splashed with liquid.

Contact Bedfont® or your local distributor for replacements:

GastroCH₄ECK® Gastrolyzer® mouthpieces

Breath bags

Sampling line

Moisture removal filter

This document is for guidance only.

Visit www.bedfont.com/resources to view this document in other languages.



Bedfont® Scientific Ltd.
Station Road, Harrietsham, Maidstone,
Kent, ME17 1JA England
Tel: +44 (0)1622 851122 Fax: +44 (0)1622 854860
Email: ask@bedfont.com Web: www.bedfont.com

© Bedfont® Scientific Limited 2024

Issue 5 - September 2024, Part No: LAB807_USA
Bedfont® Scientific Limited reserves the right to change or update this literature without prior notice.
Registered in: England and Wales. Registered No: 1289798

Rx Only



MD 502905